

## ONE REVOLUTION DIAL INDICATORS

SHOCK  
PROOF

INSPECTION  
CERTIFICATE

- Meet DIN878
- Graduation: 0.01 mm
- Eliminate the possibility of reading errors due to miscounting
- Shockproof
- Supplied with limit pointers
- Optional accessory: backs (page 277~278), contact points (page 272~276)



2316-05

| Code      | Range | Dial indicator stroke | Accuracy | Hysteresis | Dial reading | Maximum measuring force | Cap         | Remark    |
|-----------|-------|-----------------------|----------|------------|--------------|-------------------------|-------------|-----------|
| 2316-05F* | 0.5mm | 5mm                   | 9μm      | 3μm        | 25-0-25      | 1.5N                    | cannot lift | flat back |
| 2316-1F*  | 1mm   | 5mm                   | 9μm      | 3μm        | 50-0-50      | 1.5N                    | cannot lift | flat back |
| 2316-05*  | 0.5mm | 5mm                   | 9μm      | 3μm        | 25-0-25      | 1.5N                    | cannot lift | lug back  |
| 2316-1*   | 1mm   | 5mm                   | 9μm      | 3μm        | 50-0-50      | 1.5N                    | cannot lift | lug back  |

\* Supplied with manufacturer inspection certificate

